

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID: M107980
Date Received: 04/20/06
Date Extracted: 04/21/06
Date Analyzed: 04/21/06
Matrix: Water
Units: ug/L (ppb)

Client: Alaskan Copper Works
Project: Metro Self Monitor, PO# M107980
Lab ID: 604187-01 10x
Data File: 604187-01 10x.049
Instrument: ICPMS1
Operator: btb

Internal Standard:	% Recovery:	Lower	Upper
Germanium	81	Limit:	Limit:
		60	125

Analyte:	Concentration ug/L (ppb)
Chromium	1,170
Nickel	1,440
Copper	1,180
Zinc	34.4

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID: Method Blank
Date Received: Not Applicable
Date Extracted: 04/21/06
Date Analyzed: 04/21/06
Matrix: Water
Units: ug/L (ppb)

Client: Alaskan Copper Works
Project: Metro Self Monitor, PO# M107980
Lab ID: I6-224 mb
Data File: I6-224 mb.041
Instrument: ICPMS1
Operator: btb

Internal Standard:	% Recovery:	Lower Limit:	Upper Limit:
Germanium	98	60	125

Analyte:	Concentration ug/L (ppb)
Chromium	<10
Nickel	<10
Copper	<10
Zinc	<10

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 05/03/06

Date Received: 04/20/06

Project: Metro Self Monitor, PO# M107980, F&BI 604187

QUALITY ASSURANCE RESULTS FOR THE ANALYSIS OF WATER SAMPLES FOR METALS BY EPA METHOD 200.8

Laboratory Code: 604120-03 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	ug/L (ppb)	34.5	34.9	1	0-20
Nickel	ug/L (ppb)	62.1	59.7	5	0-20
Copper	ug/L (ppb)	26.4	25.1	5	0-20
Zinc	ug/L (ppb)	61.0	73.9	19	0-20

Laboratory Code: 604120-03 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	% Recovery MS	Acceptance Criteria
Chromium	ug/L (ppb)	20	34.5	92 b	50-150
Nickel	ug/L (ppb)	20	62.1	75 b	50-150
Copper	ug/L (ppb)	20	26.4	85 b	50-150
Zinc	ug/L (ppb)	50	61.0	92 b	50-150

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	% Recovery LCS	Acceptance Criteria
Chromium	ug/L (ppb)	20	103	70-130
Nickel	ug/L (ppb)	20	101	70-130
Copper	ug/L (ppb)	20	104	70-130
Zinc	ug/L (ppb)	50	96	70-130

b - The analyte was spiked at a level that was less than five times that present in the sample. Matrix spike recoveries may not be meaningful.

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

May 3, 2006

DUPLICATE COPY

INVOICE # 06ACU0503-2

Accounts Payable
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

RE: Project Metro Self Monitor, PO# M107980, F&BI 604187 - Results of testing
requested by Gerry Thompson for material submitted on April 20, 2006.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc by Method 200.8 @ \$80 per sample	<u>\$ 80.00</u>
Amount Due	\$ 80.00

FEDERAL TAX ID # (b) (6)

604187

SAMPLE CHAIN OF CUSTODY

ME 04/20/06

AI 4

Send Report To

G. Thompson

Company

ALASKAN Copper works

Address

628 S. Harbor St

City, State, ZIP

Seattle WA 98134

Phone #

206-382-8379

Fax #

206-382-4309

SAMPLERS (Signature)

PROJECT NAME/NO.

METRO Self monitor

PO #

M107980

REMARKS

Page # of

TURNAROUND TIME

☒ Standard (2 Weeks)☐ RUSH

Rush charges authorized by:

SAMPLE DISPOSAL

☐ Dispose after 30 days☐ Return samples☐ Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS	For Crypton				
M107980	01	4/29/06	11:00	H2O	1											

Friedman & Bruya, Inc.
3012 16th Avenue West

Seattle, WA 98119-2029

Ph. (206) 285-8282

Fax (206) 283-5044

FORMS\COC\COC.DOC

SIGNATURE

Relinquished by:

Received by:

Relinquished by:

Received by:

PRINT NAME

COMPANY

DATE

TIME

G. Thompson

Nhan Phan

ALASKA Copper

FBI

4/20/06

4/20/06

1:48 PM

1:48

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

May 3, 2006

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

Dear Mr. Thompson:

Included are the results from the testing of material submitted on April 20, 2006 from the Metro Self Monitor, PO# M107980, F&BI 604187 project. There are 3 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Michael Erdahl
Project Manager

Enclosures
ACU0503R.DOC